


- e-CHECK\* (XS)** Quality Control for the **ProCyte Dx\*** Analyzer
- e-CHECK\* (XS)** contrôle qualité pour l'analyseur **ProCyte Dx\***
- e-CHECK\* (XS)** Control de calidad para el analizador **ProCyte Dx\***
- e-CHECK\* (XS)** Qualitätskontrolle für das **ProCyte Dx\*** Analysegerät
- e-CHECK\* (XS)** Controllo qualita' per l'analizzatore **ProCyte Dx\***
- e-CHECK\* (XS)** IDEXXプロサイトDx\*の精度管理


**L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)**

**LOT** 40780804

 06/09/24

**L2 (Normal / Normale / 正常濃度)**

**LOT** 40780805

 06/09/24

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 <sup>9</sup> /μL)	2.31	2.16 - 2.46
HCT (%)	15.8	12.8 - 18.8
HGB (g/dL)	5.2	4.9 - 5.5
MCV/VGM (fL)	68.6	61.1 - 76.1
MCH/TCMH (pg)	22.5	18.7 - 26.3
MCHC/CCMH (g/dL)	32.9	21.7 - 44.1
PLT (K/μL)	55	25 - 85
RDW-SD (fL)	48.1	39.1 - 57.1
RDW-CV (%)	20.2	16.7 - 23.7
PDW (fL)	9.3	5.3 - 13.3
PCT (%)	0.05	0.03 - 0.07
MPV (fL)	8.1	5.1 - 11.1
P-LCR (%)	9.5	2.0 - 17.0
WBC/GB (K/μL)	3.05	2.65 - 3.45
NEUT# (K/μL)	1.33	0.92 - 1.74
LYMPH# (K/μL)	1.09	0.69 - 1.49
MONO# (K <sup>9</sup> /μL)	0.35	0.05 - 0.65
EO# (K/μL)	0.28	0.08 - 0.48
RBC-O (10 <sup>9</sup> /μL)	2.34	1.94 - 2.74
PLT-O (K/μL)	46	16 - 76
RETIC (K/μL)	157.4	117.4 - 197.4
% RETIC (%)	6.81	5.11 - 8.51

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 <sup>9</sup> /μL)	4.39	4.12 - 4.66
HCT (%)	35.6	32.1 - 39.1
HGB (g/dL)	12.0	11.5 - 12.5
MCV/VGM (fL)	81.1	73.1 - 89.1
MCH/TCMH (pg)	27.3	23.6 - 31.0
MCHC/CCMH (g/dL)	33.7	27.6 - 39.8
PLT (K/μL)	217	172 - 262
RDW-SD (fL)	46.6	39.6 - 53.6
RDW-CV (%)	17.1	14.1 - 20.1
PDW (fL)	9.0	6.0 - 12.0
PCT (%)	0.18	0.10 - 0.26
MPV (fL)	8.5	5.5 - 11.5
P-LCR (%)	9.5	3.6 - 15.4
WBC/GB (K/μL)	7.35	6.60 - 8.10
NEUT# (K/μL)	3.60	2.54 - 4.66
LYMPH# (K/μL)	2.20	1.58 - 2.82
MONO# (K <sup>9</sup> /μL)	0.82	0.06 - 1.58
EO# (K/μL)	0.73	0.37 - 1.09
RBC-O (10 <sup>9</sup> /μL)	4.26	3.56 - 4.96
PLT-O (K/μL)	183	133 - 233
RETIC (K/μL)	122.3	92.3 - 152.3
% RETIC (%)	2.79	1.99 - 3.59